

Report By:

National TAB  
1329 E. KEMPER ROAD  
SUITE 4210  
CINCINNATI, OH 45246



**Report: Hotworx (Belmont, CA) TAB REPORT**

**Function: Test, Adjust, & Balance**

**Date: 02/19/2024**

# PROJECT

## Hotworx (Belmont, CA)

410 El Camino Real

Belmont, CA 94002

### Client

Martinico & Sons, Inc.

1776 S. 7th St.

San Jose, CA 95112

# National TAB

Project: Hotworx (Belmont, CA)

## Table Of Contents

<b>Section</b>	<b>Page #</b>
Certification	3
Equipment Calibrations	4
Abbreviations	7
GRD Layout	8
Fan Coil	10
FAN - Exhaust	14



# CERTIFICATION

**PROJECT:** Hotworx (Belmont, CA)

The data presented in this report is a record of system measurements and final adjustments that have been obtained in accordance with the current edition of the NEBB *Procedural Standards for Testing, Adjusting, and Balancing of Environmental Systems*. Any variances from design quantities, which exceed NEBB tolerances, are noted in the Test-Adjust-Balance Report Project Summary.

The air distribution system has been tested and balanced and final adjustments have been made in accordance with NEBB standards and the project specifications.

**NEBB TAB FIRM:** National TAB-Southeast

**REGISTRATION NO:** 3755

**CERTIFIED BY:** J. Scott Springer 23312

**DATE:** 2/19/2024

The hydronic distribution system has been tested and balanced and final adjustments have been made in accordance with NEBB standards and the project specifications.

**NEBB TAB FIRM:** National TAB-Southeast

**REGISTRATION NO:** 3086


**CERTIFIED BY:** J. Scott Springer 23312

**DATE:** \_\_\_\_\_

## Submitted and Certified by:

**NEBB TAB FIRM:** National TAB-Southeast

**TAB PROFESSIONAL:** J. Scott Springer

**SIGNATURE:** 

**REGISTRATION NO:** 3755 (NTAB) / 23312

**CERTIFICATION EXP:** 12/31/2024





# National TAB

## Testing, Adjusting, and Balancing Equipment



Function		Range	Minimum Accuracy	Instrument Information	Calibration Date	Date Due
AIR	AIR PRESSURE	0 in wg to 10 in wg	2% +/- 0.001 in wg	TSI EBT731 EBT732117009	9/7/2023	9/7/2024
	AIR VELOCITY INSTRUMENT	50 fpm to 3900 fpm	+/- 5 % +/- 7 fpm	TSI EBT731 EBT732117009	9/7/2023	9/7/2024
	DIRECT HOOD READING	100 cfm to 2000 cfm	+/- 5 % +/- 7 cfm	TSI EBT731 EBT732117009	9/7/2023	9/7/2024
TEMPERATURE	AIR METER	-20 F to 240 F	+/- .5 % 2 F	Cooper SRH77A S/N 100516003	9/29/2023	9/29/2024
	AIR PROBE	-20 F to 240 F	+/- .5 % 2 F	Cooper SRH77A S/N 100516003	9/29/2023	9/29/2024
	IMMERSION METER	-20 F to 240 F	+/- .5 % 2 F	Cooper SRH77A S/N 100516003	9/29/2023	9/29/2024
	IMMERSION PROBE	-20 F to 240 F	+/- .5 % 2 F	Cooper SRH77A S/N 100516003	9/29/2023	9/29/2024
	CONTACT METER	-20 F to 240 F	+/- .5 % 2 F	Cooper SRH77A S/N 100516003	9/29/2023	9/29/2024
	CONTACT PROBE	-20 F to 240 F	+/- .5 % 2 F	Cooper SRH77A S/N 100516003	9/29/2023	9/29/2024
HUMIDITY	HUMIDITY PROBE	10 % RH to 90 % RH	3% of reading	Cooper SRH77A S/N 100516003	9/29/2023	9/29/2024
ELECTRICAL	VOLTAGE MEASUREMENT	0 VAC to 600 VAC	2 % reading +/- 5 digits	Klein Tools CL800 S/N 1220C-C1	9/29/2023	9/29/2024
	AMPERAGE MEASUREMENT	0 Amperes to 100 Amperes	2 % reading +/- 5 digits	Klein Tools CL800 S/N 1220C-C1	9/29/2023	9/29/2024
ROTATION	ROTATION MEASUREMENT	60 rpm to 5000 rpm	2 % reading 2 rpm	Shimpo DT 207Lp S/N D1690029R	9/29/2023	9/29/2024



# National TAB

Testing, Adjusting, and Balancing Equipment



### Report of Calibration

**Kansas City Calibration Lab., Inc.**  
8847 Long Street  
Lenexa, Kansas 66215

Telephone: (913) 541-0629 Internet: [www.kccl.com](http://www.kccl.com) Email: [service@kccl.com](mailto:service@kccl.com)

UNIT UNDER TEST: TSI EBT731 Differential Digital Meter	TEST RESULT: PASS
SERIAL NUMBER: EBT732117009	PERFORMED ON: 9/7/2023
ASSET NUMBER: EBT732117009	DATA TYPE: FOUND-LEFT
PROCEDURE NAME: ADM-XXX / EBT-XXX-XX 2.0% Reading: 1 Yr Cert CPC	TEMPERATURE: 23.8°C
PROCEDURE REV.: 20210930C	HUMIDITY: 44 %
CALIBRATED BY: Bart Schwartz	BAROMETRIC: 28.93 inHg
P.O. NUMBER:	<b>Recalibration Date</b> September 07, 2024
CUSTOMER: National TAB 1126 Swift Street NKC, MO 64116	Calibration Number: 0007333
Cal Seals Intact: Yes	Previous Calibration Date: August 12, 2022

K.C. Calibration Lab., Inc. certifies that the above listed instrument meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). This calibration is traceable to the International System of Units (SI), through National Metrology Institutes (NIST, PTB NRC NPL, etc), radiometric techniques, or natural physical constants. This calibration complies with MIL-STD-45662A and ANSI/NCSL Z540-1-1994.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the calibration organization issuing this report.

Note: Any Test Uncertainty Ratio (TUR) that is less than four to one will appear under the "TUR" heading on the data record. If the TUR meets or exceeds four to one, the field is left blank.

REMARKS:

Asset#	Description	Cal Date	Due Date
41001AR6	Mensor CPC6050 Low & Medium Pressure Calibrator	3/15/2023	3/15/2024

Test Description	True Value	Test Result	Lower Limit	Upper Limit	Units	TUR
Vertical Accuracy: ±2.0%±0.001 Reading INH20 @ 60"						
Reference Documents: Mfr. Manual						
Internal Barometric Reading: 28.80 inHg						
Version Number is 1.11.1						
** Connector						
0.000 inH2O	0.000	-0.0005	-0.0100	0.0100	inH2O	Pass
5.000 inH2O	5.000	5.1000	4.9000	5.1000	inH2O	Pass
10.000 inH2O	10.000	10.0300	9.8000	10.2000	inH2O	Pass
14.900 inH2O	14.900	14.9100	14.6020	15.1980	inH2O	Pass
0.000 inH2O	0.000	-0.0003	-0.0100	0.0100	inH2O	Pass
-5.000 inH2O	-5.000	-5.0100	-5.1000	-4.9000	inH2O	Pass
-10.000 inH2O	-10.000	-10.0200	-10.2000	-9.8000	inH2O	Pass
-14.900 inH2O	-14.900	-14.9600	-15.1980	-14.6020	inH2O	Pass

Report of Calibration for SERIAL NUMBER: EBT732117009 ASSET NUMBER: EBT732117009

Printed On: Thursday, September 7, 2023 Page 1 of 2  
**Test Results** indicate the following: Found-Left: Unit was left as found. As-Left: Unit was left after adjustments.

Test Description	True Value	Test Result	Lower Limit	Upper Limit	Units	TUR
** Connector						
0.000 inH2O	0.000	-0.0001	-0.0100	0.0100	inH2O	Pass
5.000 inH2O	5.000	4.9800	4.9000	5.1000	inH2O	Pass
10.000 inH2O	10.000	10.0300	9.8000	10.2000	inH2O	Pass
14.900 inH2O	14.900	14.9100	14.6020	15.1980	inH2O	Pass
0.000 inH2O	0.000	0.0001	-0.0100	0.0100	inH2O	Pass
-5.000 inH2O	-5.000	-5.0100	-5.1000	-4.9000	inH2O	Pass
-10.000 inH2O	-10.000	-10.0300	-10.2000	-9.8000	inH2O	Pass
-14.900 inH2O	-14.900	-14.9200	-15.1980	-14.6020	inH2O	Pass

\*\*\*\*\*END OF CALIBRATION\*\*\*\*\*

K.C. Calibration Labs Seal

Signature: *Bart Schwartz*  
Bart A. Schwartz, Engineer in Charge

Report of Calibration for SERIAL NUMBER: EBT732117009 ASSET NUMBER: EBT732117009

Printed On: Thursday, September 7, 2023 Page 2 of 2  
**Test Results** indicate the following: Found-Left: Unit was left as found. As-Left: Unit was left after adjustments.

### Report of Calibration

**Kansas City Calibration Lab., Inc.**  
8847 Long Street  
Lenexa, Kansas 66215

Telephone: (913) 541-0629 Internet: [www.kccl.com](http://www.kccl.com) Email: [service@kccl.com](mailto:service@kccl.com)

UNIT UNDER TEST: Shimpo DT-2077p Tachometer	TEST RESULT: PASS
SERIAL NUMBER: D1690029R	PERFORMED ON: 9/29/2023
ASSET NUMBER: D1690029R	DATA TYPE: FOUND-LEFT
PROCEDURE NAME: Shimpo DT-20xx: 1 Year Certification	TEMPERATURE: 24.9°C
PROCEDURE REV.: 20210818C	HUMIDITY: 47 %
CALIBRATED BY: Bart Schwartz	<b>Recalibration Date</b> September 29, 2024
P.O. NUMBER:	Calibration Number: 00077544
CUSTOMER: National TAB 1126 Swift Street NKC, MO 64116	Previous Calibration Date: August 11, 2022
Cal Seals Intact: Yes	

K.C. Calibration Lab., Inc. certifies that the above listed instrument meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). This calibration is traceable to the International System of Units (SI), through National Metrology Institutes (NIST, PTB NRC NPL, etc), radiometric techniques, or natural physical constants. This calibration complies with MIL-STD-45662A and ANSI/NCSL Z540-1-1994.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the calibration organization issuing this report.

Note: Any Test Uncertainty Ratio (TUR) that is less than four to one will appear under the "TUR" heading on the data record. If the TUR meets or exceeds four to one, the field is left blank.

REMARKS:

Asset#	Description	Cal Date	Due Date
MYS900813	Keysight Technologies 33511B Function/Arb Waveform Generator	12/1/2022	12/1/2023

Test Description	True Value	Test Result	Lower Limit	Upper Limit	Units	TUR
RPM						
10.00 RPM	10.0	10	9	11	RPM	Pass
100.00 RPM	100.0	100	99	101	RPM	Pass
1000.0 RPM	1000.0	1000	999	1001	RPM	Pass
10,000.0 RPM	10000.0	10000	9998	10002	RPM	Pass
99,900.0 RPM	99900.0	99902	99893	99907	RPM	Pass

Report of Calibration for SERIAL NUMBER: D1690029R ASSET NUMBER: D1690029R

Printed On: Friday, September 29, 2023 Page 1 of 2  
**Test Results** indicate the following: Found-Left: Unit was left as found. As-Left: Unit was left after adjustments.

### Report of Calibration

**Kansas City Calibration Lab., Inc.**  
8847 Long Street  
Lenexa, Kansas 66215

Telephone: (913) 541-0629 Internet: [www.kccl.com](http://www.kccl.com) Email: [service@kccl.com](mailto:service@kccl.com)

UNIT UNDER TEST: Cooper Instrument SRH77A Digital Thermometer	TEST RESULT: PASS
SERIAL NUMBER: 100516003	PERFORMED ON: 9/29/2023
ASSET NUMBER: 100516003	DATA TYPE: FOUND-LEFT
PROCEDURE NAME: Met Temp NIST(SI) 1 Year	TEMPERATURE: 24.1°C
PROCEDURE REV.:	HUMIDITY: 46 %
CALIBRATED BY: Bart Schwartz	<b>Recalibration Date</b> September 29, 2024
P.O. NUMBER:	Calibration Number: 00077543
CUSTOMER: National TAB 1126 Swift Street NKC, MO 64116	Previous Calibration Date: August 12, 2022
Cal Seals Intact: Yes	

K.C. Calibration Lab., Inc. certifies that the above listed instrument meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). This calibration is traceable to the International System of Units (SI), through National Metrology Institutes (NIST, PTB NRC NPL, etc), radiometric techniques, or natural physical constants. This calibration complies with MIL-STD-45662A and ANSI/NCSL Z540-1-1994.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the calibration organization issuing this report.

Note: Any Test Uncertainty Ratio (TUR) that is less than four to one will appear under the "TUR" heading on the data record. If the TUR meets or exceeds four to one, the field is left blank.

REMARKS:

Asset#	Description	Cal Date	Due Date
2659119	Hart Scientific 1523 Single Chan Reference Thermometer	1/9/2023	1/9/2024
905040	Burns Engineering 5615 Platinum Resistance Thermometer	2/8/2023	2/8/2024
DWS18	Fluke 518 Dry-Block Calibrator	8/28/2023	8/28/2024
MB7103	Hart Scientific 7103 Micro Bath Calibrator	12/8/2022	12/8/2023

Test Description	True Value	Test Result	Lower Limit	Upper Limit	Units	TUR
1075 General Purpose Puncture Probe						
Accuracy ±1.3 deg F / ±0.2 deg C or ±0.5% of reading:						
-10.00 F	-10.08	-8.70	1.38			
32.00 F	32.34	32.70	0.36			
122.00 F	122.71	121.80	-0.91			
212.00 F	211.90	211.10	-0.80			
280.00 F	279.96	280.70	0.74			
4011 Pipe Strap Probe						
Accuracy ±2% Range -25° to 212°F / -32° to 100°C						
0.00 F	0.27	2.10	1.83			
75.00 F	75.25	75.10	-0.15			
150.00 F	150.31	150.00	-0.31			

Report of Calibration for SERIAL NUMBER: 100516003 ASSET NUMBER: 100516003

Printed On: Friday, September 29, 2023 Page 1 of 2  
**Test Results** indicate the following: Found-Left: Unit was left as found. As-Left: Unit was left after adjustments.



# National TAB

Testing, Adjusting, and Balancing Equipment



Test Results	True Value	Test Result	Lower Limit	Upper Limit	Units	TUR
5028 Slim Humidity Probe						
Accuracy ±2% from 20 to 80%RH, ±3% below 20 and ±						
10.0 %RH @ 23.0°C	%RH	10.0	16	6.0		
25.0 %RH @ 23.0°C	%RH	25.0	30	5.0		
50.0 %RH @ 23.0°C	%RH	50.0	53	3.0		
75.0 %RH @ 23.0°C	%RH	75.0	77	2.0		
23.0°C @ 10.0 %RH	C	23.0	23.2	0.2		
23.0°C @ 25.0 %RH	C	23.0	23.2	0.2		
23.0°C @ 50.0 %RH	C	23.0	23.2	0.2		
23.0°C @ 75.0 %RH	C	23.0	23.1	0.1		

\*\*\*\*\*END OF CALIBRATION\*\*\*\*\*

Signed: *Bart A. Schwartz*  
Bart A. Schwartz, Engineer in Charge

Report of Calibration for SERIAL NUMBER: 100516003 ASSET NUMBER: 100516003 Page 2 of 2

Printed On: Friday, September 29, 2023  
Test Results indicate the following: Found-Left: Unit was left as found. As-Left: Unit was left after adjustments.

### Report of Calibration

**Kansas City Calibration Lab., Inc.**  
8847 Long Street  
Lenexa, Kansas 66215

Telephone: (913) 541-0629 Internet: [www.kccl.com](http://www.kccl.com) Email: [service@kccl.com](mailto:service@kccl.com)

UNIT UNDER TEST: Klein Tools CL800 True RMS Digital Clampmeter	TEST RESULT: PASS
SERIAL NUMBER: 1220C-C1	PERFORMED ON: 9/29/2023
ASSET NUMBER: 1220C-C1	DATA TYPE: FOUND-LEFT
PROCEDURE NAME: Klein Tools CL800 : (1 year) CAL VER / 5520	TEMPERATURE: 24.9°C
PROCEDURE REV.: 20230928	HUMIDITY: 46%
CALIBRATED BY: Bart Schwartz	
P.O. NUMBER:	
CUSTOMER: National TAB 1126 Swift Street NKC, MO 64116	Recalibration Date September 29, 2024 Calibration Number: 0007542 Previous Calibration Date:

Cal Seals Intact: Unknown

K.C. Calibration Lab., Inc. certifies that the above listed instrument meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). This calibration is traceable to the International System of Units (SI), through National Metrology Institutes (NIST, PTB, NRC, NPL, etc), radiometric techniques, or natural physical constants. This calibration complies with MIL-STD-45662A and ANSI/NCISL Z540-1-1994.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the calibration organization issuing this report.

Note: Any Test Uncertainty Ratio (TUR) that is less than four to one will appear under the "TUR" heading on the data record. If the TUR meets or exceeds four to one, the field is left blank.

REMARKS:

Standards Used	Asset #	Description	Cal Date	Due Date
	3277903	Fluke 5522A Multi-Product Calibrator	11/30/2022	11/30/2023

Test Results	True Value	Test Result	Lower Limit	Upper Limit	Units	TUR
Root Difference Square guardbanding method used						
AC VOLTS TESTS						
6 V Range						
5.900 V @ 60 Hz	5.9000	5.897	5.807	5.994	V	Pass
60 V Range						
59.00 V @ 60 Hz	59.0000	58.97	58.24	59.76	V	Pass
600 V Range						
590.0 V @ 60 Hz	590.000	589.7	582.4	597.6	V	Pass
1000 V Range						
990.0 V @ 60 Hz	990.000	991.0	970.1	1009.9	V	Pass
DC VOLTS TESTS						
600 mV Range						
600.0 mV	600.000	599.8	593.2	606.8	m V	Pass

Report of Calibration for SERIAL NUMBER: 1220C-C1 ASSET NUMBER: 1220C-C1

Printed On: Friday, September 29, 2023  
Test Results indicate the following: Found-Left: Unit was left as found. As-Left: Unit was left after adjustments.

Test Results	True Value	Test Result	Lower Limit	Upper Limit	Units	TUR
-600.0 mV						
	-600.00	-599.9	-606.8	-593.2	m V	Pass
6 V Range						
6.000 V	6.0000	5.997	5.937	6.063	V	Pass
-6.000 V	-6.0000	-5.995	-6.063	-5.937	V	Pass
60 V Range						
60.00 V	60.0000	59.96	59.37	60.63	V	Pass
600 V Range						
600.0 V	600.0000	599.6	593.7	606.3	V	Pass
1000 V Range						
1000.0 V	1000.0000	1001	985	1015	V	Pass
-1000.0 V	-1000.0000	-1001	-1015	-985	V	Pass
CONTINUITY TESTS						
Audible Indicator ON @ 10 ohms						
Audible Indicator OFF @ 51 ohms						
RESISTANCE TESTS						
600 Ohm Range						
600.0 Ohm	600.000	601.2	590.5	609.5	Ω	Pass
6 kOhm Range						
6.000 kOhm	6.00000	6.000	5.905	6.095	k Ω	Pass
60 kOhm Range						
60.00 kOhm	60.0000	59.99	59.05	60.95	k Ω	Pass
600 kOhm Range						
600.0 kOhm	600.0000	599.9	590.5	609.5	k Ω	Pass
6 MOhm Range						
6.000 MOhm	6.00000	5.993	5.905	6.095	M Ω	Pass
60 MOhm Range						
60.00 MOhm	60.00000	59.47	58.70	61.30	M Ω	Pass
DIODE CHECK TESTS						
Diode Voltage						
FREQUENCY TESTS						
9.00 Hz @ 8 V						
9.00 Hz @ 8 V	9.0000	8.999	8.905	9.095	Hz	Pass
90.00 Hz @ 8 V						
90.00 Hz @ 8 V	90.0000	89.95	89.05	90.95	Hz	Pass
900.0 Hz @ 8 V						
900.0 Hz @ 8 V	900.0000	900.0	890.5	909.5	Hz	Pass
9.000 kHz @ 8 V						
9.000 kHz @ 8 V	9.00000	9.000	8.905	9.095	k Hz	Pass
90.00 kHz @ 8 V						
90.00 kHz @ 8 V	90.00000	90.00	89.05	90.95	k Hz	Pass
100.0 kHz @ 8 V						
100.0 kHz @ 8 V	100.00000	100.00	98.5	101.5	k Hz	Pass
DUTY CYCLE						
50.0 % @ 1 kHz						
50.0 % @ 1 kHz	50.000	50.3	49.3	50.8	%	Pass
CAPACITANCE TESTS						
60 nF Range						
60.00 nF	60.0000	59.96	55.70	62.30	n F	Pass
600 nF Range						
600.0 nF	600.0000	597.1	571.8	608.2	n F	Pass
6 uF Range						
6.000 uF	6.00000	5.854	5.718	6.082	u F	Pass
60 uF Range						
60.00 uF	60.00000	58.87	57.18	60.82	u F	Pass

Report of Calibration for SERIAL NUMBER: 1220C-C1 ASSET NUMBER: 1220C-C1

Printed On: Friday, September 29, 2023  
Test Results indicate the following: Found-Left: Unit was left as found. As-Left: Unit was left after adjustments.

Test Results	True Value	Test Result	Lower Limit	Upper Limit	Units	TUR
6000 uF Range						
6000 uF	6000.000	5996.6	5660.0	6200.0	u F	Pass
TEMPERATURE F TESTS						
5900 uF						
5900 uF	5900.000	5957	5600	6200	u F	Pass
-14 °F						
-14 °F	-14.0	-10	-23	-5	°F	Pass
100 °F						
100 °F	100.0	102	94	106	°F	Pass
500 °F						
500 °F	500.0	502	490	510	°F	Pass
900 °F						
900 °F	900.0	902	873	927	°F	Pass
TEMPERATURE C TESTS						
-25 °C						
-25 °C	-25.0	-23	-31	-20	°C	Pass
100 °C						
100 °C	100.0	102	96	104	°C	Pass
350 °C						
350 °C	350.0	351	344	357	°C	Pass
500 °C						
500 °C	500.0	501	485	515	°C	Pass
AC CURRENT TESTS						
60 A Range						
50.00 A @ 60 Hz	50.0000	49.60	48.92	51.08	A	Pass
50.00 A @ 400 Hz	50.0000	50.00	48.92	51.08	A	Pass
400 A Range						
500.0 A @ 60 Hz	500.0000	494.2	489.5	510.5	A	Pass
500.0 A @ 100 Hz	500.0000	494.4	489.5	510.5	A	Pass
60 A Range						
50.00 A	50.0000	49.20	48.92	51.08	A	Pass
600 A Range						
300.0 A	300.0000	296.5	293.5	306.5	A	Pass
590.0 A	590.0000	582.7	577.7	602.3	A	Pass

\*\*\*\*\*END OF CALIBRATION\*\*\*\*\*

Signed: *Bart A. Schwartz*  
Bart A. Schwartz, Engineer in Charge


Report of Calibration for SERIAL NUMBER: 1220C-C1 ASSET NUMBER: 1220C-C1


Printed On: Friday, September 29, 2023  
Test Results indicate the following: Found-Left: Unit was left as found. As-Left: Unit was left after adjustments.



## Abbreviation List

A = Area (ft <sup>2</sup> )	S.F. = Service Factor
AHU = Air Handling Unit	SF = Supply Fan
A <sub>k</sub> = Effective Area	SP = Static Pressure
BHP = Brake Horsepower (IP) HP	SR = Supply Register
Btu = British Thermal Unit	T = Temperature
Btu/h = Btuh = BTUH = BTU/Hour	T <sub>ma</sub> = Mixed Air Temperature
CL = Center Distance (used in belt formula)	T <sub>oa</sub> = Outside Air Temperature
CD = Ceiling Diffuser	T <sub>ra</sub> = Return Air Temperature
CF = Correction Factor	H = Head (in wc, ft wc, psi)
CFM = Volumetric Flow: Cubic Feet Per Minute	h = Enthalpy
CO <sub>2</sub> = Carbon Dioxide	HP = Horsepower
CO = Carbon Monoxide	hr = Hour
C <sub>v</sub> = Flow Constant	K <sub>v</sub> = Flow constant (SI)
d = Diameter (in.) IP	kW = Kilowatt = 1000 Watts
Δ = Difference or Change (Final - Initial)	LAT = Leaving Air Temperature
DB = Dry Bulb	lb = Pounds
EA = Exhaust Air	LWT = Leaving Water Temperature
EAT = Entering Air Temperature	ma = Mixed Air
EF = Exhaust Fan	MIN = Minimum
Eff = Efficiency	MAX = Maximum
EG = Exhaust Grille	N/A = Not Applicable
ESP = External Static Pressure	NA = No Access
EWT = Entering Water Temperature	NL = Not Listed
°F = Degrees Fahrenheit, °F	NPSHA = Net Positive Suction Head Available
FPB = Fan Powered Box	NS = Not Specified
FLA = Full Load Amps	OA = Outside Air
fpm = Feet per Minute (fpm)	OAT = Outside Air Temperature
ft = Foot	PD = Sheave Pitch Diameter
gal = Gallons	P.D. = Pressure Drop
GPM = Gallons Per Minute (GPM)	PF = Power Factor
h = Enthalpy (BTU/lb dry air)	SG = Supply Grille
P = Pressure	SR = Supply Register
ppm = parts per million	TP = Total Pressure
psi = Pounds Per Square Inch	T <sub>ra</sub> = Return Air Temperature
psid = PSI Differential	TS = Tip Speed (fpm) IP, (m/s) SI
r = Radius (in)	TSP = Total Static Pressure
% <sub>ra</sub> = % of Return Air	V = Velocity
RA = Return Air	VAV = Variable Air Volume
RAT = Return Air Temperature	VD = Volume Damper
RF = Return Fan	VFD = Variable Frequency Drive
RG = Return Grille	W = Watt
RH = Relative Humidity	WB = Wet Bulb
RPM = Revolutions Per Minute	wg = wc = water gauge = water column
RTU = Roof Top Unit	WHP = Water Horsepower (IP)
SA = Supply Air	ω = Humidity Ratio





# National TAB

Project: Hotworx (Belmont, CA)

System/Unit: Fan Coil



Asset: FC1

AREA:EXERCISE SPACE

Unit Data		
	Design	Actual
MFG	NA	BRYANT
Model Num	NA	524FP07H000A2UAAAA
Serial Num	-	4023U18244
Configuration	HORIZONTAL	HORIZONTAL

Motor Data		
	Design	Actual
Horsepower	-	2.4
Phase	3	3
Voltage (rated)	208	208
Amperage (rated)	-	6.4

Test Data		
	Design	Actual
SFAN CFM	2870	3070
Motor Speed SetPt	-	A
RL Voltage	-	214/214/215
RL Amperage	-	1.25/1.25/1.26
RA CFM	2370	2559
OA CFM	500	511

Performance Data		
	Design	Actual
Suction ESP	-	-0.15"
Discharge ESP	-	0.28"
Total ESP	0.1	0.43"

Completed By: Zack Eismin on 02/16/2024

# National TAB

Project: Hotworx (Belmont, CA)

## Fan Coil



### Diffuser Supply (GRD)

#### FC1/EXERCISE SPACE

Asset							
Asset Name	Location	Type	Size	DESIGN CFM	CFM(1)	FINAL CFM	% to design
FC1-SGRD1	EXERCISE	SD1	18X6	210	439	221	105.2
FC1-SGRD2	SAUNA #5	SD2	38X14	1330	2927	1411	106.1
FC1-SGRD3	SAUNA #4	SD2	38X14	1330	3271	1438	108.1
Total				2870	6637	3070	106.97%

# National TAB

Project: Hotworx (Belmont, CA)

## System/Unit: Fan Coil



Asset: FC2

AREA:EXERCISE SPACE

Unit Data		
	Design	Actual
MFG	NA	BRYANT
Model Num	NA	524FP07H000A2UAAAA
Serial Num	-	4023U18245
Configuration	HORIZONTAL	HORIZONTAL

Motor Data		
	Design	Actual
Horsepower	-	2.4
Phase	3	3
Voltage (rated)	208	208
Amperage (rated)	-	6.4

Test Data		
	Design	Actual
SFAN CFM	3000	3159
Motor Speed SetPt	-	A
RL Voltage	-	214/215/214
RL Amperage	-	1.3/1.4/1.3
RA CFM	2440	2586
OA CFM	560	573

Performance Data		
	Design	Actual
Suction ESP	-	-0.14"
Discharge ESP	-	0.31"
Total ESP	0.1	0.45"

Completed By: Zack Eismin on 02/16/2024

# National TAB

Project: Hotworx (Belmont, CA)

## Fan Coil



### Diffuser Supply (GRD)

#### FC2/EXERCISE SPACE

Asset							
Asset Name	Location	Type	Size	DESIGN CFM	CFM(1)	FINAL CFM	% to design
FC2-SGRD1	HALL	SD2	38X14	1330	3116	1438	108.1
FC2-SGRD2	SAUNA #1	SD2	38X14	1330	3008	1388	104.4
FC2-SGRD3	UTILITY	CD1	6	60	130	61	101.7
FC2-SGRD4	LOBBY	CD1	8	140	350	135	96.4
FC2-SGRD5	LOBBY	CD1	8	140	340	137	97.9
Total				3000	6944	3159	105.3%

# National TAB

Project: Hotworx (Belmont, CA)

## System/Unit: FAN - Exhaust



Asset: EF1

AREA:ELEC RM

Unit Data		
	Design	Actual
MFG	NA	PANASONIC
Model Num	NA	FV-1115VK2
Serial Num	-	31106M
Type	CEILING	CEILING

Test Data		
	Design	Actual
CFM	112	107
RL Voltage	-	NA
RL Amperage	-	NA
Total ESP	0.25	0.19"

Motor Data		
	Design	Actual
Motor MFG	-	NL
Frame	-	NL
Horsepower	-	NL
Motor Rpm	-	NL
Phase	1	1
Voltage (rated)	120	120
Amperage (rated)	-	0.33
Service Factor	-	NL

Completed By: Zack Eismin on 02/16/2024

# National TAB

Project: Hotworx (Belmont, CA)

## System/Unit: FAN - Exhaust



Asset: EF2

AREA:RESTROOM

Unit Data		
	Design	Actual
MFG	NA	PANASONIC
Model Num	NA	FV-0511VF1
Serial Num	-	31121
Type	CEILING	CEILING

Test Data		
	Design	Actual
CFM	80	86
RL Voltage	-	NA
RL Amperage	-	NA
Total ESP	0.25	NA

Motor Data		
	Design	Actual
Motor MFG	-	NL
Frame	-	NL
Horsepower	-	NL
Motor Rpm	-	NL
Phase	1	1
Voltage (rated)	120	120
Amperage (rated)	-	0.29
Service Factor	-	NL

Completed By: Zack Eismin on 02/16/2024

# National TAB

Project: Hotworx (Belmont, CA)

## System/Unit: FAN - Exhaust



Asset: EF3

AREA:RESTROOM

Unit Data		
	Design	Actual
MFG	NA	PANASONIC
Model Num	NA	FV-0411VF1
Serial Num	-	31121
Type	CEILING	CEILING

Test Data		
	Design	Actual
CFM	80	83
RL Voltage	-	NA
RL Amperage	-	NA
Total ESP	0.25	NA

Motor Data		
	Design	Actual
Motor MFG	-	NL
Frame	-	NL
Horsepower	-	NL
Motor Rpm	-	NL
Phase	1	1
Voltage (rated)	120	120
Amperage (rated)	-	0.29
Service Factor	-	NL

Completed By: Zack Eismin on 02/16/2024

# National TAB

Project: Hotworx (Belmont, CA)

## System/Unit: FAN - Exhaust



Asset: EF4

AREA:UTILITY RM

Unit Data		
	Design	Actual
MFG	NA	PANASONIC
Model Num	NA	FV-1115VK2
Serial Num	-	31121
Type	CEILING	CEILING

Test Data		
	Design	Actual
CFM	112	110
RL Voltage	-	NA
RL Amperage	-	NA
Total ESP	0.25	0.18"

Motor Data		
	Design	Actual
Motor MFG	-	NL
Frame	-	NL
Horsepower	-	NL
Motor Rpm	-	NL
Phase	1	1
Voltage (rated)	120	120
Amperage (rated)	-	0.33
Service Factor	-	NL

Completed By: Zack Eismin on 02/16/2024

# National TAB

Project: Hotworx (Belmont, CA)

## System/Unit: FAN - Exhaust



Asset: EF5

AREA:EXERCISE SPACE

Unit Data		
	Design	Actual
MFG	NA	GREENHECK
Model Num	NA	CUBE-100-4-1-19-X
Serial Num	-	23644621
Type	CRE	UPBLAST

Test Data		
	Design	Actual
CFM	600	572
RL Voltage	-	NA
RL Amperage	-	NA
Total ESP	0.50	0.57"

Motor Data		
	Design	Actual
Motor MFG	-	WEG
Frame	-	56
Horsepower	-	0.25
Motor Rpm	-	1765
Phase	3	3
Voltage (rated)	208	208
Amperage (rated)	-	1.07
Service Factor	-	1.35

Completed By: Zack Eismin on 02/16/2024

# National TAB

Project: Hotworx (Belmont, CA)

## FAN - Exhaust



Diffuser Ret/Exh (GRD)

### EF5/EXERCISE SPACE

Asset								
Asset Name	Type	Size	DESIGN CFM	AK	CFM(1)	CFM(2)	FINAL CFM	% to design
EF5-EGRD1	EG1	18X10	600	1	521	572	572	95.3
Total			600		521	572	572	95.33%

Completed By: Zack Eismin on 02/16/2024

# National TAB

Project: Hotworx (Belmont, CA)

## System/Unit: FAN - Exhaust



Asset: EF6

AREA:SAUNA #3

Unit Data		
	Design	Actual
MFG	NA	GREENHECK
Model Num	NA	CUE-070-6-1-17-X
Serial Num	-	23644629
Type	CRE	UPBLAST

Test Data		
	Design	Actual
CFM	150	142
RL Voltage	-	NA
RL Amperage	-	NA
Total ESP	0.25	0.22"

Motor Data		
	Design	Actual
Motor MFG	-	GREENHECK
Frame	-	NL
Horsepower	-	1/30
Motor Rpm	-	1550
Phase	1	1
Voltage (rated)	115	115
Amperage (rated)	-	1.1
Service Factor	-	NL

Completed By: Zack Eismin on 02/16/2024

# National TAB

Project: Hotworx (Belmont, CA)

## FAN - Exhaust



Diffuser Ret/Exh (GRD)

### EF6/SAUNA #3

Asset								
Asset Name	Type	Size	DESIGN CFM	AK	CFM(1)	CFM(2)	FINAL CFM	% to design
EF6-EGRD1	EG2	10X6	150	0.37	142	142	142	94.7
Total			150		142	142	142	94.67%